

# **Correlation spectroscopy of solid microparticles suspended in aqueous medium**

Autor(en): **Lorusso, G.F. / Minafra, A. / Capozzi, V.**

Objekttyp: **Article**

Zeitschrift: **Helvetica Physica Acta**

Band (Jahr): **62 (1989)**

Heft 6-7

PDF erstellt am: **27.04.2024**

Persistenter Link: <https://doi.org/10.5169/seals-116153>

## **Nutzungsbedingungen**

Die ETH-Bibliothek ist Anbieterin der digitalisierten Zeitschriften. Sie besitzt keine Urheberrechte an den Inhalten der Zeitschriften. Die Rechte liegen in der Regel bei den Herausgebern.

Die auf der Plattform e-periodica veröffentlichten Dokumente stehen für nicht-kommerzielle Zwecke in Lehre und Forschung sowie für die private Nutzung frei zur Verfügung. Einzelne Dateien oder Ausdrucke aus diesem Angebot können zusammen mit diesen Nutzungsbedingungen und den korrekten Herkunftsbezeichnungen weitergegeben werden.

Das Veröffentlichen von Bildern in Print- und Online-Publikationen ist nur mit vorheriger Genehmigung der Rechteinhaber erlaubt. Die systematische Speicherung von Teilen des elektronischen Angebots auf anderen Servern bedarf ebenfalls des schriftlichen Einverständnisses der Rechteinhaber.

## **Haftungsausschluss**

Alle Angaben erfolgen ohne Gewähr für Vollständigkeit oder Richtigkeit. Es wird keine Haftung übernommen für Schäden durch die Verwendung von Informationen aus diesem Online-Angebot oder durch das Fehlen von Informationen. Dies gilt auch für Inhalte Dritter, die über dieses Angebot zugänglich sind.

Ein Dienst der *ETH-Bibliothek*

ETH Zürich, Rämistrasse 101, 8092 Zürich, Schweiz, [www.library.ethz.ch](http://www.library.ethz.ch)

## CORRELATION SPECTROSCOPY OF SOLID MICROPARTICLES SUSPENDED IN ACQUEOUS MEDIUM

**G.F. Lorusso, A. Minafra, V. Capozzi, D. Loiacono**

Dipartimento di Fisica, Universita' di Bari and Unita'

CISM-GNSM di Bari, Via Amendola 163, 70126 Bari, Italy

To study the optical and morphological properties of solid microparticles, the technique of the intensity fluctuation spectroscopy can be used [1,2]. A correlator set up employing a personal computer has been carried out. This apparatus is able to analyze the intensity fluctuation of the scattered light by microparticles suspension, at room temperature, illuminated by a He-Ne laser. We have used Polystirene spheres sizing from 0.1 to 0.2  $\mu\text{m}$  for the calibration of the correlation system. From the analysis of the correlation function it is possible to obtain the size distribution of suspended solid particles.

The theory of intensity fluctuation can be carried out by using a model developed by Einstein and Smoluchowski in 1908, which assumes an homogeneous medium with dielectric constant fluctuation

$$\epsilon(\mathbf{r}, t) = \epsilon_0 \mathbf{I} + \delta\epsilon(\mathbf{r}, t)$$

The theory gives scattered electric field and scattered intensities. So, information on the medium inhomogeneity (i.e.  $\delta\epsilon$ ) can be achieved experimentally by measuring the spectral density of the scattered light  $I_s(\mathbf{q}, \omega)$ , or, alternatively, by analyzing the fluctuation of integrated scattered intensity  $I_s(\mathbf{q})$ . This last approach is developed by measuring the intensity correlation function  $g^{(2)}$ , which is connected to the field correlation function  $g^{(1)}$  by the Siegert relation

$$g^{(2)} = (g^{(1)})^2 + 1$$

Note that  $g^{(1)}$  is the Fourier transform of the spectral density.

In the particular case of spherical particles suspended in a liquid  $g^{(1)}$  is given by

$$g^{(1)} = \exp[-q^2 D t]$$



FIG.1 Correlation function  $g^{(2)}$ 

FIG.2 Histogram analysis

where  $q = 4\pi(n/\lambda)\sin\theta/2$  and  $D = (k_B T_m)/(6\pi\eta a)$ , in which  $n$  is the medium refractive index,  $\lambda$  the laser wavelength,  $\theta$  the angle of the scattered photons,  $a$  the particle radius,  $T_m$  the medium temperature and  $\eta$  its viscosity. The intensity correlation function  $g^{(2)}$  can be obtained by measuring the photon counting correlation function

$$g^{(2)} = \frac{\langle n(0, T)n(\tau, T) \rangle}{\langle n \rangle^2}$$

where  $T$  is the sampling time and  $\tau$  is the delay time. By dividing the measuring time in  $r$  channels of duration  $T$ ,  $g(2)$  is written as

$$g_T^{(2)}(\tau T) = \frac{N \sum_N n_T(\tau T)n_T(0)}{[\sum_N n_T(0)]^2}$$

The correlation function  $g^{(2)}$  obtained for 215 nm nominal diameter particles is shown in fig.1, where  $T$  is  $100 \mu s$ . The analysis has made on  $6 \cdot 10^5$  counts. The obtained diameter is  $d=209.38$  nm,  $\Delta d/d=0.02$ .

Determination of size distribution for particles of different diameters requires an analysis of experimental results using a convolution integral of the type

$$g^{(1)} = \int G(\Gamma) \exp(-\Gamma t) d\Gamma$$

where  $G(\Gamma)$  is the size distribution of particles.  $G(\Gamma)$  can be approximated by various methods; one of the most effective is the histogram one [3]. The software necessary for the histogram analysis has been developed and tested by using a computer simulated size distribution. A typical result is shown in fig.2.

- [1] Berne B.J., Pecora R. "Dynamic Light Scattering", Wiley and Sons
- [2] Chu B. "Laser Light Scattering", Academic Press
- [3] Fletcher G.C., Ramsay D.J., Optica Acta, 1983, 30, 8, 1183